



CAP Commonly Cited Deficiencies and How to Avoid Them

Objectives

- Review the most cited deficiencies in 2023
- Explain how to improve laboratory processes to prevent deficiencies
- Examine CAP inspection tools and accreditation resources



CAP Top Ten Deficiencies


Top 10 Deficiencies in 2023

Checklist Requirement		CAP Ranking # of Time Cited
COM.10000	Policy and Procedure Manual	#1, 741
COM.01200	Activity Menu	#2, 712
COM.04250	Comparability of Instruments and Methods Nonwaived Testing	#3, 694
GEN.55500	Competency Assessment Non-waived Testing	#4, 644
COM.01700	PT and Alternative Assessment Result Evaluation	#5, 610
COM.30600	Maintenance/Function Checks	#6, 598
COM.04200	Instrument/Equipment Record Review	#7, 504
COM.01100	Ungraded PT Challenges	#8, 425
COM.30750	Temperature Checks	#9, 421
COM.01400	PT Attestation Statement	#10, 402

Policy and Procedure Manual COM.10000


- Complete procedure manual is available:
 - Paper-based
 - Electronic
 - Web-based format

Procedures must match practice.




Poll Question #1

- In your facility, what form of procedure manual do you use?
 - Paper-based
 - Electronic
 - Both



Common Deficiencies and How to Avoid Them – Procedure Manual

- Practice does not match procedure.
- Procedures are not available at the bench level.
- Staff are unaware of how to locate electronic procedures.




Activity Menu COM.01200

- Laboratory's current CAP Activity Menu accurately reflects the testing performed.
 - Add to new test implementation process.
 - Audit Activity Menu periodically.
 - Remove retired tests.
 - Custom checklist generated by Activity Menu selections.



Comparability of Instruments and Methods COM.04250

- More than one nonwaived instrument/method to test for a given analyte.
- Instruments and methods are checked at least twice a year.



Comparability of Instruments and Methods (continued)

- Non-waived methods only
- Methods within a single CAP/CLIA number
- At least twice per year
- Applies to instruments/methods producing the same reportable results
- Written procedures including acceptance criteria

Common Deficiencies and How to Avoid Them – Comparability of Instruments and Methods

- Missing documentation of two times per year
- Missing acceptability criteria
- Does not include all non-waived testing



Competency Assessment GEN.55500

The competency of personnel performing nonwaived testing is assessed at the required frequency at the laboratory (CAP/US-based CLIA number) where testing is performed.

- All variations must be included.
- May be maintained centrally within a healthcare system but must be available upon request.



Competency Assessment (continued)

During the first year of an individual's duties, competency must be assessed at least semiannually and annually thereafter.

- Prior to performing patient testing, training must be completed and evaluated for proper test performance.
- Training and competency assessments are separate processes.
- Applicable to new testing personnel only.



Competency Assessment (continued)

Assessment includes the applicable six elements of competency noted under GEN.55500 for each test system.

- Use laboratory activity menu to identify test systems.
 - Same analyte with two test systems (eg, automated, manual) needs separate competency assessments.
 - Multiple analytes under single test system do not need separate competency assessments (eg, chemistry panel).
 - Each test system includes assessment of Pre-analytic, Analytic, Post-analytic steps in the testing process.

Competency Assessment (continued)

The six elements of competency include:

- 1. Direct observations of routine patient test performance
- 2. Monitoring the recording and reporting of test results
- 3. Review of intermediate test results or worksheets, quality control records, proficiency testing results, and preventive maintenance records
- 4. Direct observation of performance of instrument maintenance and function checks
- 5. Assessment of test performance through testing previously analyzed specimens or proficiency testing specimens
- **6.** Evaluation of problem-solving skills

Common Deficiencies and How to Avoid Them – Competency Assessments

- Missing all 6 elements of competency
- Competency events are not traceable

Competency Assessment - Example

Ele-	Specify Instrument				
ments	/ Assay	Chemistry Analyzer	LC-TOF	GC-FI	GC-MS
	,	, ,			
1	Patient ID/Prep	n/a	n/a	n/a	n/a
1	Specimen Collection	n/a	n/a	n/a	n/a
1	Handling/Processing	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
		01/08/18 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
1	Testing	Accession # M123456	Accession # M123456	Accession # M123456	Accession # M123456
		01/08/18 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
2	Reporting Criticals	Accession # M123456	Accession # M123456	Accession # M123456	Accession # M123456
		01/08/18 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
2	Reporting Normals	Accession # M123456	Accession # M123456	Accession # M123456	Accession # M123456
3	Review worksheets	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
3	Review QC	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
		03/15/18 SLM	03/15/18 SLM	03/15/18 SLM	03/15/18 SLM
3	Review PT results	Sample UDS-15	Sample UDS-16	Sample UDC-16	Sample UNK-17
		0.410.014.0.01.14	0.410.014.0.01.14	0.4/0.0/4.0 01.14	0.4/0.0/4.0 01.14
3	Review PM records	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
	••	04/00/45 CLM	04/00/40 CLM	04/00/40 CLM	04/00/40 CLM
4	Maintenance	01/08/15 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
_	B	02/17/18 SLM	02/15/18 SLM	02/15/18 SLM	02/15/18 SLM
5	Proficiency Testing	Sample UDS-15	Sample UDS-16	Sample UDC-16	Sample UNK-17
_	Dlind Commiss	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM	01/08/18 SLM
5	Blind Samples	Accession # M234567	Accession # M234567	Accession # M234567	Accession # M234567
	Buchlana Octobra	Written Quiz = 100%	Trouble Shooting Log	Abnormal diff quiz =	Verbal quiz = 100%
6	Problem Solving	01/08/18 SLM	01/09/18 SLM	100%	01/08/18 SLM

PT and Alternative Assessment Evaluation COM.01700

- Ongoing evaluation of PT/EQA and alternative assessment results
- Corrective action taken for each unacceptable result
 - Any result or sample not meeting defined acceptability criteria must be evaluated
 - Investigate for impact on patient sample result
 - Correction of problems appropriate to the failure are performed in a timely manner.

Common Deficiencies and How to Avoid Them – PT/APA Evaluation

- Missing corrective actions on failures
- Missing documentation of review of results with codes
- Missing documentation or evaluation of alternative assessments



PT/EQA Exception Investigation Worksheet




PT Exception Investigation Worksheet

investigation workshee						leet		
Survey Information								
Survey Name:			CAP No.					
Date Survey Received:			Date Analysis Performed:					
Date Survey Results Subn	nitted:		Date Results Received:					
Investigation Performed B	y:							
Analyte:							_	
Specimen Number	Reported Result	Intended	Result/Range	Acceptable/	Inaccenta	hle	7	
Specimen Number	Reported Result	IIICIIGGG	Result Range	Acceptable	Diraccepta	ibic	-	
							┪	
							_	
	_						\dashv	
Evaluation of Possible S	ources of Error							
Clerical					YES	NO	N/A	
Were the results submitted	d by the due date?							
Was the result correctly tra	anscribed from the inst	rument read	l-out or report?					
Was the correct instrumen	t/method/reagent repo	rted on the	result form?					
Do the units of measure m	atch between the resu	ılt form and	the instrument result	s?				
Is the decimal place correct?								
Does the result reported or evaluation report?	n the result form matcl	h the result t	found on the proficie	ncy testing				
A response of "No" to any is unlike those for patient r provided with the proficien testing device. If results re contact your proficiency te	results, clerical errors r cy testing, addition of ported on the result for	nay indicate a second re	a need for additiona	al staff training, re on of the reportir	eview of in	struction	ns	


Maintenance/Function Checks COM.30600

- Appropriate maintenance and function checks are performed
- Records retained following a defined schedule
 - All instruments and equipment
 - Written procedure
 - Schedule specified by manufacturer
 - Documentation of performance and monthly review




Common Deficiencies and How to Avoid Them – Maintenance/Function Checks

- No documentation of required preventive maintenance
- Missing documentation of maintenance
- No corrective actions for missed maintenance



Instrument/Equipment Record Review COM.04200

- Documentation must be reviewed and assessed at least monthly
 - Laboratory Director review
 - Designee review



Common Deficiencies and How to Avoid Them – Instrument/Equipment Record Review

- Missing documentation
- Missing acceptability criteria
- Does not include all non-waived testing



Instrument/Equipment Review Example

			Fill in the date the document review occurred for that month										
Department	Instrument/Testing	Jan	Feb	Mar	Apr	May	Jun	Jul	Aug	Sep	Oct	Nov	Dec
	Room Temperature Logs	02/06/22											
Lab	Refrigerator Temperature Logs	02/06/22											
₽	Freezer Temperature Logs	02/06/22											
	Eye wash Logs / Shower Logs	02/06/22											
	Instrument A maintenance logs	02/15/22											
	Instrument A QC logs	02/15/22											
	Instrument A calibration logs	><	><		><	\geq					><	$\geq <$	$\geq <$
	Instrument B maintenance logs	02/15/22											
stry	Instrument B QC logs	02/15/22											
i E	Instrument B calibration logs	><	$\geq <$		><	$\geq \leq$	$\geq \leq$				><		$\geq <$
ું ક	Instrument A & B Comparisons	><	$\geq <$	$\geq <$		\geq				$\geq <$		><	$\geq <$
	Blood Gas maintenance logs	02/15/22											
	Blood Gas QC logs	02/15/22											
	Blood Gas calibration logs	02/15/22	><	><	><	><	><		><	><	><	$\geq <$	><
	PT Records	02/27/22											

Ungraded PT Challenges COM.01100

- Laboratory director or designee assesses performance of PT challenges that are ungraded
 - Records must show ungraded PT results are evaluated for acceptable performance
 - Investigation
 - Corrective action

Common Deficiencies and How to Avoid Them – Ungraded PT Challenges

- Laboratory not defining how it assesses performance on ungraded PT challenges
- Missing signatures or notation of acceptability
- Failing to retain records of review

Temperature Checks COM.30750

Temperatures are checked and recorded for all temperature-dependent equipment and environments

- Can use min/max thermometers
- Corrective actions when temperatures are out of range



Poll Question #2

- In your facility, which temperature checks do you use?
 - Manual
 - Electronic
 - Both



Common Deficiencies and How to Avoid Them – Temperature Checks

- Missing documentation of corrective actions when temperatures are out
- Temperature ranges are not set for all items/materials with the area
- Missing documentation of weekend monitoring if closed

COM.01400 PT Attestation Statement

- PT/EQA attestation statement is signed by:
 - Laboratory director or designee
 - All individuals involved in the testing process
 - Physical or secured electronic signatures must be present

Attestation/Use of Other Form							
Attestation Statement							
As stated in the February 28, 1992 United States Federal Register under Subpart H 493-801 (b) (1), "the individual testing or examining the samples and the laboratory director must attest to the routine integration of the samples into the patient work load using the laboratory's routine methods." The laboratory director or designee and the testing personnel must sign on the result form.							
You may use the attestation page provided in the kit instructions or, alternatively, print, sign, and retain a copy of this page for your records and inspection purposes.							
If your laboratory requires additional space f	or signatures, copy this form as needed.						
We, the undersigned, recognizing that some special handling may be required due to the nature of proficiency testing (PT) materials, have as closely as is practical, performed the analyses on these specimens in the same manner as regular patient specimens. We confirm that results were not shared or PT specimens referred or tested outside our CLIA identification number.							
Director (or Designee) (signature required)		Survey Mailing Information					
<u>010</u>	<u>010</u> 070						
							
040							
Testing Personnel (signature required) Testing Personnel (signature required) Testing Personnel (signature required)							
080	110	140					

Common Deficiencies and How to Avoid Them – PT/EQA Attestation Statement

- Missing signature
- Transfusion Medicine or other blood bank-related PT/EQA signed by unqualified personnel


Attestation/Use of Other Form						
Attestation Statement						
As stated in the February 28, 1992 United States Federal Register under Subpart H 493-801 (b) (1), "the individual testing or examining the samples and the laboratory director must attest to the routine integration of the samples into the patient work load using the laboratory's routine methods." The laboratory director or designee and the testing personnel must sign on the result form.						
You may use the attestation page provided in the kit instructions or, alternatively, print, sign, and retain a copy of this page for your records and inspection purposes.						
If your laboratory requires additional space for signatures, copy this form as needed.						
We, the undersigned, recognizing that some s	pecial handling may be required due to the nature o					
We, the undersigned, recognizing that some s closely as is practical, performed the analyses not shared or PT specimens referred or tested Director (or Designee) (signature required)	pecial handling may be required due to the nature o					
We, the undersigned, recognizing that some significance is closely as is practical, performed the analyses not shared or PT specimens referred or tested	pecial handling may be required due to the nature o	patient specimens. We confirm that results were Survey Mailing Information				
We, the undersigned, recognizing that some siclosely as is practical, performed the analyses not shared or PT specimens referred or tested Director (or Designee) (signature required)	pecial handling may be required due to the nature o	patient specimens. We confirm that results were Survey Mailing Information				

Top 10 Deficiencies in 2023

Checklist Requirement		CAP Ranking # of Time Cited
COM.10000	Policy and Procedure Manual	#1, 741
COM.01200	Activity Menu	#2, 712
COM.04250	Comparability of Instruments and Methods Nonwaived Testing	#3, 694
GEN.55500	Competency Assessment Non-waived Testing	#4, 644
COM.01700	PT and Alternative Assessment Result Evaluation	#5, 610
COM.30600	Maintenance/Function Checks	#6, 598
COM.04200	Instrument/Equipment Record Review	#7, 504
COM.01100	Ungraded PT Challenges	#8, 425
COM.30750	Temperature Checks	#9, 421
COM.01400	PT Attestation Statement	#10, 402

CAP Resources to Keep Up-to-Date

- CAP Today
- e-Alerts
- Online Inspector Training –
 Team Member/Team Leader
- CAP Accreditation Resources Repository
- Educational webinars –
 Focus on Compliance Series





Newly Expanded Accreditation Resources: CAP Accredited Laboratories

- Revised and expanded online resources make it easier to find the answers you seek.
- New content includes:
 - A series of Checklist Q&A's written by technical specialists
 - An informative multi-module course, *Laboratory Inspection Preparation:*
 - Getting Ready for Your First Inspection
- Everything is fully searchable to find what you need quickly.

CAP's e-LAB Solutions Suite is available at any time for accreditation questions.

Questions?

